



SMITH RIVER BIOLOGICALS

Quality Biologicals And Prepared Media For The Laboratory Professional

Certificate of Analysis

Violet Red Bile Agar with MUG

Product Number: 11-VRBAM

Fill Volume: 20mL

Lot Number: 1908070

Expiration: 11/7/2019

This is to certify that representative samples of this lot of media were tested in accordance with routine microbiological procedures in our quality control laboratory. The following organisms were used and gave the following results and reactions:

Performance Testing

<u>Test Organisms</u>	<u>Growth</u>	<u>Reaction</u>
<i>E. coli</i> ATCC 25922	4+	Pink
<i>E. faecalis</i> ATCC 29212	Inhibited	N/A
<i>S. typhimurium</i> ATCC 14028	4+	Pink
<i>E. aerogenes</i> ATCC 13048	4+	Clear

Microbial testing was performed aerobically at 35°C for 18-24 hours

Growth approved by: _____
(QA Manager)

Microbial Load

Report is based on U.S.P.XXI(1211)

Lot Size: 2L

Number of Units Sampled: 27

Final pH at 25°C: 7.4±0.2

Appearance

Uninoculated media is pale to medium purple and clear, may contain dark spots

Lot approved by: _____
(QA Manager)

THIS DOCUMENT MEETS THE STANDARDS FOR INSPECTION AND ACCREDITATION CURRENTLY ESTABLISHED BY THE COLLEGE OF AMERICAN PATHOLOGIST, NCCLS-M22T, U.S.P.XXI(1211) AND MS105D. NOT INTENDED FOR THE DIAGNOSIS OR TREATMENT OF HUMAN DISEASE.